



TO ELIMINATE FALSE AND UNWANTED CALLS, PULL-CORD EMERGENCY STATIONS SHOULD NOT BE LOCATED NEAR TOWEL BARS OR GRAB RAJLS IN BATH ROOMS.

NOTE: ALL DIMENSIONS ARE TO CENTER OF UNIT.

FLOOR LEVEL

RECOMMENDED MOUNTING HEIGHTS

SCALE: N.T.S.	APPROVED BY:	DRAWN BY: D.J.G.
DATE: JAN. 10, '78		REVISED:

VERIFY ACTUAL HT. V. ARCHT. DWG.



CORNELL ELECTRONIC PRODUCTS, INC.
MILWAUKEE, WISCONSIN

DRAWING NUMBER
011078-1

This drawing, data and design herein shall not be duplicated, used, or disclosed in whole or in part for any other purpose, except as otherwise authorized by contract, without written permission of Cornell Electronic Products, Inc. All responsibilities shall bear this notice.

NO.	DATE	REVISIONS